江苏省计量科学研究院

JIANGSU INSTITUTE OF METROLOGY

校 准 证 书

Calibration Certificate

证 书 编 号：
Certificate No. H2019-0003638

客 户 名 称 Customer	江苏省精创电气股份有限公司
客 户 地 址 Customer Add.	江苏省徐州铜山经济开发区黄山路 1 号
计量器具名称 Name of Instrument	无线温湿度监测仪
型 号 / 规 格 Type/Specification	ECW-50
出 厂 编 号 Serial No.	37317 56201
制 造 单 位 Manufacturer	Elitech

（校准专用章）
Stamp

批 准 人 Approved by	栾海峰	职 务 Position	所总工
校 准 员 Calibrated by	谢业雷	核 验 员 Checked by	王建雷

校准日期 Calibration Date：2019 年 Year 1 月 Month 22 日 Day

计量检定机构授权证书号：（国）法计（2017）01022
Authorization Certificate No.(2017)01022
地址：南京市栖霞区文澜路 95 号（总部）
Add: No.95,Wenlan Road,Qixia District,Nanjing（Headquarter）
电话：(025) 84636990
Tel
电子信箱：jsimguest@jsim.com.cn
E-mail
网址：www.jsim.com.cn
Website
邮编：210023
Post Code
传真：(025)84636972
Fax

本院是国家法定计量检定机构，计量授权证书号：（国）法计（2017）01022 号。

This laboratory is a national legal metrological verification institute. Authorization certificate No. (2017)01022.

本次校准项目已通过中国合格评定国家认可委员会能力认可。实验室认可证书号：CNAS L0841。

The calibration ability of this item has been assessed by China National Accreditation Service for Conformity Assessment (CNAS).

Certificate of laboratory accreditation: No. CNAS L0841.

本院出具的数据均可溯源至国家计量基准。

All data issued by the laboratory are traceable to national measurement standards.

本次校准的技术依据：

Reference documents for the calibration

JJF 1366-2012 《温度数据采集仪校准规范》

本次校准所使用的主要计量标准器具：

Main measurement standards used in the calibration

名称/编号 Name/Number	测量范围 Measuring Range	准确度等级/最大允许误差/不确定度 Class of Accuracy/Maximum Permissible Errors/ Uncertainty of Measurement	证书编号/有效期至 Certificate No./Valid to
标准铂电阻温度计/03722	(-189.3442～+419.527)℃	一等	RGcp2017-0199/19-04-10

校准地点、环境条件：

Place and environmental conditions of the calibration

地点： 南京市栖霞区文澜路 95 号 7321 室　　温度： 22℃　　相对湿度： 50%

Place No.95,Wenlan Road,Qixia District,Nanjing Room 7321　　Temperature　　R. Humidity

校准结果/说明见续页

Results of calibration and additional explanation in next page.

注： 1、未经本院书面授权，不得部分复制本证书。

Note This certificate shall not be reproduced except in full, without the written approval of the laboratory.

2、本证书的校准结果仅对本次被校计量器具有效。

The results are only responsible for the items calibrated.

3、本证书未加盖校准专用章无效。

This certificate will be invalid if it is issued without official stamp.

校准结果/说明（续页）

Results of calibration and additional explanation (continued page)

校准点(℃)	误差值(℃)	扩展不确定度（U）(k=2)（℃）
-20.0	+0.3	0.1
0.0	0.0	0.1
40.0	+0.2	0.1

以下空白。

JSIM 江苏计量 Jiangsu Metrology